

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3075

Name: RITTER2  
Sterilizer ID: V2313228  
11/02/2023 08:37 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.4 F 0.0  
2:00 71.7 F 0.0  
4:00 83.6 F 0.3  
6:00 117.3 F 1.5  
8:00 151.8 F 4.5  
10:00 171.0 F 8.4  
12:00 193.3 F 12.5  
14:00 222.1 F 5.6  
16:00 223.7 F 9.4  
18:00 235.6 F 13.6  
20:00 232.9 F 9.5  
22:00 237.9 F 9.9  
24:00 245.2 F 13.0  
26:00 257.4 F 19.6  
28:00 268.3 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.2  
1:00 271.4 F 29.3  
2:00 271.5 F 29.4  
3:00 271.4 F 29.4  
4:00 271.4 F 29.2

Min 271.4 F 29.1  
Max 273.0 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.8 F 0.0  
5:00 235.7 F 0.0  
10:00 206.4 F 0.0  
15:00 184.7 F 0.0  
20:00 181.2 F 0.0  
25:00 178.5 F 0.0  
30:00 177.3 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 29:54  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:07:56  
CYCLE COMPLETE

11/02/2023 09:44 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_