

-----  
Midmark M11 - v1.0.5  
Total Cycles: 967

Name: RITTER2  
Sterilizer ID: V2313228  
04/07/2022 09:20 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 178.9 F 0.0  
2:00 176.8 F 1.1  
4:00 203.9 F 6.4  
6:00 222.8 F 12.1  
8:00 227.9 F 8.3  
10:00 238.0 F 11.2  
12:00 243.8 F 12.9  
14:00 243.8 F 12.6  
16:00 257.5 F 19.8  
18:00 269.1 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.8  
1:00 272.6 F 29.4  
2:00 271.2 F 28.8  
3:00 272.7 F 29.5  
4:00 271.5 F 28.8  
  
Min 271.1 F 28.7  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.3 F 0.0  
5:00 199.6 F 0.0  
10:00 166.5 F 0.0  
15:00 164.1 F 0.0  
20:00 158.7 F 0.0  
25:00 156.8 F 0.0  
30:00 155.5 F 0.0  
Drying Complete

FILLING: 1:44  
HEATING: 19:01  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:57:55  
CYCLE COMPLETE

04/07/2022 10:41 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02