
Midmark M11 - v1.0.0
Total Cycles: 5202

Name: RITTER 1
Sterilizer ID: V2178198
12/18/2024 01:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.8 F 0.0
2:00 149.0 F 0.0
4:00 166.1 F 0.1
6:00 193.4 F 0.8
8:00 210.1 F 4.6
10:00 222.1 F 8.5
12:00 231.9 F 12.5
14:00 233.4 F 8.3
16:00 244.8 F 13.3
18:00 242.6 F 11.5
20:00 240.1 F 10.1
22:00 253.0 F 16.5
24:00 264.0 F 23.0
26:00 272.1 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 27.9
1:00 271.7 F 28.1
2:00 272.6 F 28.7
3:00 272.2 F 28.3
4:00 271.6 F 27.9

Min 271.4 F 27.8
Max 272.6 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.5 F 0.0
5:00 190.1 F 0.0
10:00 189.9 F 0.0
15:00 199.1 F 0.0
20:00 199.2 F 0.0
25:00 207.1 F 0.0
30:00 200.0 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 26:05
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:04:39
CYCLE COMPLETE

12/18/2024 02:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____