

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5112

Name: RITTER 1  
Sterilizer ID: V2178198  
11/20/2024 09:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.4 F 0.0  
2:00 74.0 F 0.0  
4:00 85.2 F 0.0  
6:00 118.9 F 0.0  
8:00 152.9 F 0.2  
10:00 180.8 F 0.6  
12:00 197.6 F 1.9  
14:00 212.7 F 5.0  
16:00 219.9 F 8.3  
18:00 228.6 F 11.9  
20:00 229.2 F 6.8  
22:00 240.5 F 11.8  
24:00 236.6 F 9.0  
26:00 239.8 F 9.8  
28:00 248.2 F 14.1  
30:00 260.2 F 20.7  
32:00 271.2 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.1  
1:00 272.7 F 28.8  
2:00 271.7 F 28.0  
3:00 272.3 F 28.7  
4:00 272.5 F 28.6  
  
Min 271.5 F 27.9  
Max 272.8 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.7 F 0.0  
5:00 189.0 F 0.0  
10:00 192.6 F 0.0  
15:00 187.4 F 0.0  
20:00 184.9 F 0.0  
25:00 190.0 F 0.0  
30:00 238.2 F 0.0  
  
Drying Complete

FILLING: 0:42  
HEATING: 32:43  
STERILIZING: 4:00  
VENTING: 3:29  
DRYING: 30:00  
TOTAL CYCLE: 01:10:54  
CYCLE COMPLETE

11/20/2024 11:05 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_