
Midmark M11 - v1.0.0
Total Cycles: 2636

Name: RITTER 1
Sterilizer ID: V2178198
12/05/2022 11:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.1 F 0.0
2:00 166.9 F 2.1
4:00 171.4 F 4.9
6:00 201.6 F 9.1
8:00 220.6 F 6.9
10:00 230.3 F 8.9
12:00 237.8 F 10.1
14:00 239.2 F 11.3
16:00 237.2 F 9.2
18:00 250.8 F 15.9
20:00 262.8 F 22.6
22:00 272.5 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 271.8 F 28.6
2:00 272.4 F 29.1
3:00 271.6 F 28.9
4:00 271.7 F 28.6

Min 271.4 F 28.4
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 221.6 F 0.0
5:00 233.8 F 0.0
10:00 227.8 F 0.0
15:00 231.0 F 0.0
20:00 229.9 F 0.0
25:00 230.2 F 0.0
30:00 224.8 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 22:13
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:00:15
CYCLE COMPLETE

12/05/2022 12:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____