
Midmark M11 - v1.0.5
Total Cycles: 1453

Name: RITTER2
Sterilizer ID: V2313228
09/02/2022 04:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.9 F 0.0
2:00 202.2 F 3.5
4:00 205.3 F 8.9
6:00 223.2 F 8.7
8:00 235.2 F 11.1
10:00 235.4 F 9.3
12:00 238.6 F 9.3
14:00 250.1 F 15.6
16:00 262.4 F 22.8
18:00 272.5 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.7 F 29.6
2:00 271.9 F 28.9
3:00 271.8 F 29.2
4:00 272.5 F 29.5

Min 271.4 F 28.6
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.2 F 0.0
5:00 226.4 F 0.0
10:00 235.6 F 0.0
15:00 227.0 F 0.0
20:00 226.3 F 0.0
25:00 223.2 F 0.0
30:00 213.4 F 0.0
Drying Complete

FILLING: 2:25
HEATING: 18:15
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:57:47
CYCLE COMPLETE

09/02/2022 05:48 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____