
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002438
7:57 14/10/2021

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:32
PRESSURIZING 1:35
STERILIZING 2:38
136.4 C 221kPa 2:38
136.2 C 219kPa 3:08
136.4 C 219kPa 3:38
136.3 C 221kPa 4:08
136.4 C 221kPa 4:38
136.4 C 222kPa 5:08
136.2 C 219kPa 5:38
136.2 C 220kPa 6:08
136.2 C 220kPa 6:08
VENTING 6:08
AIR DRYING 7:00
CYCLE COMPLETE 7:04

Digital Signature #
99F1E3101516591C
