
Midmark M11 - v1.0.0
Total Cycles: 2729

Name: RITTER 1
Sterilizer ID: V2178198
01/03/2023 01:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.8 F 0.0
2:00 131.4 F 0.0
4:00 145.8 F 0.4
6:00 173.5 F 3.2
8:00 205.3 F 8.2
10:00 223.8 F 10.7
12:00 230.7 F 8.9
14:00 238.0 F 10.5
16:00 243.3 F 12.1
18:00 241.0 F 10.6
20:00 255.2 F 17.8
22:00 267.3 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.0 F 28.4
2:00 272.5 F 28.8
3:00 272.7 F 28.8
4:00 272.7 F 28.8

Min 271.5 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 236.1 F 0.0
Dry Cycle Aborted

FILLING: 0:46
HEATING: 23:12
STERILIZING: 4:00
VENTING: 2:46
DRYING: 1:05
TOTAL CYCLE: 00:31:49

01/03/2023 02:30 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____