
Midmark M11 - v1.0.5
Total Cycles: 1530

Name: RITTER2
Sterilizer ID: V2313228
09/22/2022 05:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.4 F 0.0
2:00 191.8 F 4.1
4:00 209.7 F 10.2
6:00 225.1 F 6.9
8:00 237.9 F 12.6
10:00 241.7 F 12.4
12:00 242.5 F 11.9
14:00 256.5 F 19.3
16:00 268.5 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.0 F 29.2
2:00 271.7 F 29.5
3:00 271.9 F 29.1
4:00 271.8 F 29.6

Min 271.2 F 28.7
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.2 F 0.0
5:00 229.3 F 0.0
10:00 238.8 F 0.0
15:00 231.5 F 0.0
20:00 229.2 F 0.0
25:00 230.2 F 0.0
30:00 227.2 F 0.0
Drying Complete

FILLING: 2:42
HEATING: 17:06
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 00:56:50
CYCLE COMPLETE

09/22/2022 06:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____