
Midmark M11 - v1.0.0
Total Cycles: 3705

Name: RITTER 1
Sterilizer ID: V2178198
08/30/2023 11:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.0 F 0.0
2:00 182.1 F 3.7
4:00 218.3 F 8.7
6:00 223.9 F 7.5
8:00 232.4 F 9.8
10:00 233.8 F 8.2
12:00 245.2 F 13.4
14:00 244.0 F 12.3
16:00 257.1 F 19.1
18:00 268.6 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.5 F 28.8
2:00 271.3 F 28.3
3:00 272.6 F 28.9
4:00 271.4 F 28.2

Min 271.2 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.0 F 0.0
5:00 235.6 F 0.0
10:00 229.2 F 0.0
15:00 228.9 F 0.0
20:00 229.6 F 0.0
25:00 229.2 F 0.0
30:00 226.2 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 19:09
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:58:27
CYCLE COMPLETE

08/30/2023 12:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____