
Midmark M11 - v1.0.0
Total Cycles: 5377

Name: RITTER 1
Sterilizer ID: V2178198
02/28/2025 11:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.6 F 0.0
2:00 169.9 F 2.2
4:00 182.5 F 5.6
6:00 195.2 F 9.4
8:00 215.8 F 11.9
10:00 219.6 F 5.8
12:00 230.0 F 10.0
14:00 236.2 F 9.2
16:00 241.1 F 11.3
18:00 237.4 F 9.1
20:00 250.2 F 15.1
22:00 260.8 F 21.1
24:00 270.5 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.4 F 29.1
2:00 271.2 F 28.6
3:00 272.1 F 29.2
4:00 271.6 F 28.7

Min 271.2 F 28.5
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.3 F 0.0
5:00 212.5 F 0.0
10:00 226.6 F 0.0
15:00 229.5 F 0.0
20:00 220.3 F 0.0
25:00 216.4 F 0.0
30:00 233.6 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 25:43
STERILIZING: 4:00
VENTING: 4:01
DRYING: 30:00
TOTAL CYCLE: 01:04:55
CYCLE COMPLETE

02/28/2025 12:06 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____