
Midmark M11 - v1.0.0
Total Cycles: 2737

Name: RITTER 1
Sterilizer ID: V2178198
01/04/2023 03:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.5 F 0.0
2:00 182.2 F 0.0
4:00 200.7 F 3.7
6:00 226.5 F 8.9
8:00 235.9 F 10.0
10:00 238.0 F 10.8
12:00 236.5 F 8.5
14:00 239.8 F 9.7
16:00 249.4 F 14.8
18:00 262.2 F 22.1
20:00 272.2 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.1 F 28.5
2:00 272.1 F 28.8
3:00 271.9 F 28.3
4:00 272.2 F 28.8

Min 271.4 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.7 F 0.0
5:00 226.4 F 0.0
10:00 237.7 F 0.0
15:00 233.3 F 0.0
20:00 226.8 F 0.0
25:00 226.1 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 20:11
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:21
CYCLE COMPLETE

01/04/2023 04:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____