

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1330

Name: RITTER2  
Sterilizer ID: V2313228  
07/21/2022 10:27 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 161.8 F 0.0  
2:00 158.2 F 0.1  
4:00 204.3 F 2.5  
6:00 215.2 F 7.6  
8:00 230.8 F 12.6  
10:00 228.7 F 8.1  
12:00 241.5 F 13.2  
14:00 240.1 F 11.2  
16:00 238.4 F 10.1  
18:00 251.9 F 16.7  
20:00 263.2 F 23.3  
22:00 272.7 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.0  
1:00 272.7 F 29.7  
2:00 271.6 F 28.9  
3:00 272.2 F 29.8  
4:00 272.4 F 29.4

Min 271.2 F 28.8  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.5 F 0.0  
5:00 233.9 F 0.0  
10:00 221.2 F 0.0  
15:00 220.0 F 0.0  
20:00 212.0 F 0.0  
25:00 212.9 F 0.0  
30:00 207.7 F 0.0  
Drying Complete

FILLING: 1:24  
HEATING: 22:18  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:00:54  
CYCLE COMPLETE

07/21/2022 11:27 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_