
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002437
14:26 22/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:35
PRESSURIZING 2:42
STERILIZING 3:45
136.0 C 220kPa 3:45
136.2 C 223kPa 4:15
136.4 C 225kPa 4:45
136.3 C 226kPa 5:15
136.3 C 224kPa 5:45
136.4 C 226kPa 6:15
136.3 C 224kPa 6:45
136.3 C 224kPa 7:15
136.3 C 224kPa 7:15
VENTING 7:15
AIR DRYING 8:12
CYCLE COMPLETE 21:30

Digital Signature #
EAFEFA5A1188C0B6
