
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000617
12:56 30/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:13
PRESSURIZING 3:22
STERILIZING 4:33
136.0 C 227kPa 4:33
136.1 C 229kPa 5:03
136.1 C 228kPa 5:33
136.1 C 230kPa 6:03
136.4 C 233kPa 6:33
136.2 C 231kPa 7:03
136.4 C 233kPa 7:33
136.1 C 228kPa 8:03
136.1 C 228kPa 8:03
VENTING 8:03
AIR DRYING 8:57
CYCLE COMPLETE 23:57

Digital Signature #
6F4638397755F1EA
