
Midmark M11 - v1.0.5
Total Cycles: 3459

Name: RITTER2
Sterilizer ID: V2313228
03/20/2024 08:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.3 F 0.0
2:00 72.8 F 0.0
4:00 81.6 F 0.0
6:00 104.0 F 1.1
8:00 139.3 F 3.6
10:00 167.2 F 6.5
12:00 183.4 F 10.0
14:00 200.4 F 7.1
16:00 214.8 F 5.9
18:00 225.1 F 9.8
20:00 233.2 F 10.7
22:00 237.3 F 10.8
24:00 234.5 F 8.7
26:00 248.1 F 14.9
28:00 259.1 F 21.0
30:00 269.4 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.5 F 29.7
2:00 271.4 F 29.2
3:00 272.5 F 29.7
4:00 271.4 F 29.0

Min 271.3 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.5 F 0.0
5:00 228.2 F 0.0
10:00 225.9 F 0.0
15:00 220.9 F 0.0
20:00 216.1 F 0.0
25:00 220.9 F 0.0
30:00 225.1 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:05
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:09:15
CYCLE COMPLETE

03/20/2024 09:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____