
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001445
11:12 09/17/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:43
PRESSURIZING 6:32
STERILIZING 8:04
136.0 C 220kPa 8:04
136.4 C 219kPa 8:34
136.5 C 221kPa 9:04
136.4 C 223kPa 9:34
136.1 C 220kPa 10:04
136.5 C 221kPa 10:34
136.4 C 220kPa 11:04
136.4 C 223kPa 11:34
VENTING 11:34
AIR DRYING 12:37
CYCLE COMPLETE 42:37

Digital Signature #
8BCAE608BF632D6C
