
Midmark M11 - v1.0.0
Total Cycles: 2693

Name: RITTER 1
Sterilizer ID: V2178198
12/19/2022 03:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.8 F 0.0
2:00 185.4 F 3.4
4:00 195.9 F 8.2
6:00 212.5 F 9.3
8:00 228.2 F 8.5
10:00 239.4 F 11.8
12:00 242.0 F 11.9
14:00 241.1 F 10.8
16:00 254.6 F 17.8
18:00 266.6 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.3
1:00 272.8 F 29.1
2:00 272.0 F 28.4
3:00 272.5 F 29.0
4:00 272.5 F 28.8

Min 271.7 F 28.2
Max 273.0 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.4 F 0.0
5:00 221.6 F 0.0
10:00 232.0 F 0.0
15:00 234.1 F 0.0
20:00 225.6 F 0.0
25:00 223.0 F 0.0
30:00 222.5 F 0.0
Drying Complete

FILLING: 2:04
HEATING: 19:23
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:58:45
CYCLE COMPLETE

12/19/2022 04:04 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____