
Midmark M11 - v1.0.5
Total Cycles: 824

Name: RITTER2
Sterilizer ID: V2313228
02/23/2022 09:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.1 F 0.0
2:00 179.5 F 2.7
4:00 181.4 F 6.9
6:00 208.4 F 12.5
8:00 221.2 F 7.6
10:00 235.5 F 13.3
12:00 238.7 F 11.3
14:00 238.7 F 10.3
16:00 252.9 F 17.4
18:00 264.9 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.7 F 29.7
2:00 271.5 F 29.1
3:00 272.9 F 29.9
4:00 272.1 F 29.2

Min 271.5 F 29.0
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 212.4 F 0.0
10:00 220.5 F 0.0
15:00 213.3 F 0.0
20:00 211.6 F 0.0
25:00 212.3 F 0.0
30:00 210.4 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 19:48
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 00:58:22
CYCLE COMPLETE

02/23/2022 10:12 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____