
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008314
8:52 13/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:50
PRESSURIZING 2:53
STERILIZING 3:55
136.4 C 221kPa 3:55
136.4 C 224kPa 4:25
136.3 C 222kPa 4:55
136.4 C 222kPa 5:25
136.1 C 218kPa 5:55
136.4 C 222kPa 6:25
136.1 C 219kPa 6:55
136.4 C 223kPa 7:25
136.4 C 223kPa 7:25
VENTING 7:25
CYCLE ABORTED 7:57
STOP BUTTON PRESSED
NOT STERILE
END TIME 7:57

Digital Signature #
76B0C44E06B6B3D1
