
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 003265
16:40 02/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:13
PRESSURIZING 2:19
STERILIZING 3:08
136.0 C 220kPa 3:08
136.1 C 223kPa 3:38
136.2 C 224kPa 4:08
136.4 C 224kPa 4:38
136.4 C 225kPa 5:08
136.4 C 222kPa 5:38
136.4 C 222kPa 6:08
136.3 C 222kPa 6:38
136.3 C 222kPa 6:38
VENTING 6:38
AIR DRYING 7:33
CYCLE COMPLETE 7:38

Digital Signature #
0E41C0C7D562CDCA
