
Midmark M11 - v1.0.5
Total Cycles: 358

Name: RITTER2
Sterilizer ID: V2313228
09/14/2021 06:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.1 F 0.0
2:00 73.4 F 0.0
4:00 75.3 F 0.1
6:00 88.2 F 0.2
8:00 124.8 F 0.8
10:00 154.6 F 4.4
12:00 178.3 F 8.3
14:00 198.1 F 12.1
16:00 227.9 F 6.6
18:00 227.6 F 11.2
20:00 233.5 F 7.7
22:00 243.5 F 13.4
24:00 243.4 F 12.0
26:00 255.4 F 18.2
28:00 265.8 F 24.7
30:00 271.5 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.3
1:00 272.5 F 29.3
2:00 272.0 F 29.5
3:00 271.6 F 29.5
4:00 271.9 F 29.5

Min 271.5 F 29.2
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.2 F 0.0
5:00 232.3 F 0.0
10:00 226.7 F 0.0
15:00 231.1 F 0.0
20:00 227.9 F 0.0
25:00 202.7 F 0.0
Dry Cycle Aborted

FILLING: 0:58
HEATING: 30:21
STERILIZING: 4:00
VENTING: 2:56
DRYING: 29:11
TOTAL CYCLE: 01:07:26

09/14/2021 07:31 AM
ACCEPTED: X
REJECTED:
Operator: U01