

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 002515  
12:05 25/10/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:32  
PRESSURIZING 3:33  
STERILIZING 4:38  
136.4 C 222kPa 4:38  
136.3 C 224kPa 5:08  
136.4 C 224kPa 5:38  
136.4 C 222kPa 6:08  
136.4 C 225kPa 6:38  
136.4 C 223kPa 7:08  
136.4 C 221kPa 7:38  
136.4 C 224kPa 8:08  
136.4 C 224kPa 8:08  
VENTING 8:08  
AIR DRYING 9:06  
CYCLE COMPLETE 9:10  
-----

Digital Signature #  
50C3C4C01F121567  
-----