
Midmark M11 - v1.0.0
Total Cycles: 1134

Name: RITTER 1
Sterilizer ID: V2178198
08/31/2021 08:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.8 F 0.0
2:00 144.8 F 0.0
4:00 153.1 F 1.0
6:00 178.3 F 3.7
8:00 197.2 F 6.9
10:00 211.1 F 11.3
12:00 223.2 F 4.9
14:00 233.2 F 9.4
16:00 241.2 F 11.2
18:00 240.7 F 11.1
20:00 237.1 F 9.0
22:00 250.6 F 15.5
24:00 261.9 F 22.0
26:00 272.5 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.7 F 29.0
2:00 271.7 F 28.3
3:00 271.7 F 28.8
4:00 272.7 F 28.9

Min 271.1 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.9 F 0.0
5:00 209.1 F 0.0
10:00 199.0 F 0.0
15:00 195.8 F 0.0
20:00 192.9 F 0.0
25:00 188.4 F 0.0
30:00 187.2 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 26:28
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:04:49
CYCLE COMPLETE

08/31/2021 09:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____