
Midmark M11 - v1.0.0
Total Cycles: 4598

Name: RITTER 1
Sterilizer ID: V2178198
05/30/2024 02:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 113.3 F 0.0
2:00 114.6 F 1.3
4:00 133.6 F 4.0
6:00 162.5 F 7.4
8:00 190.2 F 11.4
10:00 207.3 F 3.6
12:00 217.8 F 7.8
14:00 228.9 F 12.1
16:00 231.4 F 7.8
18:00 243.1 F 13.0
20:00 240.9 F 10.8
22:00 253.0 F 16.9
24:00 263.8 F 23.2
26:00 271.9 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.6 F 29.1
2:00 271.7 F 28.3
3:00 271.9 F 28.8
4:00 272.4 F 28.9

Min 271.2 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.9 F 0.0
5:00 210.4 F 0.0
10:00 219.2 F 0.0
15:00 230.9 F 0.0
20:00 228.3 F 0.0
25:00 230.0 F 0.0
30:00 230.1 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 26:11
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:04:25
CYCLE COMPLETE

05/30/2024 03:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____