
Midmark M11 - v1.0.5
Total Cycles: 2718

Name: RITTER2
Sterilizer ID: V2313228
08/01/2023 11:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.1 F 0.0
2:00 195.3 F 1.6
4:00 210.0 F 7.1
6:00 236.3 F 12.7
8:00 232.4 F 9.4
10:00 237.3 F 9.3
12:00 243.5 F 12.4
14:00 245.3 F 13.2
16:00 258.1 F 20.2
18:00 269.4 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.5 F 29.3
2:00 272.0 F 29.5
3:00 272.1 F 29.1
4:00 272.3 F 29.7

Min 271.4 F 28.8
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.0 F 0.0
5:00 236.1 F 0.0
10:00 230.1 F 0.0
15:00 217.9 F 0.0
20:00 203.3 F 0.0
25:00 200.5 F 0.0
30:00 198.0 F 0.0
Drying Complete

FILLING: 2:15
HEATING: 18:59
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:28
CYCLE COMPLETE

08/01/2023 12:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____