
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001493
10:34 10/13/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 3:18
STOP BUTTON PRESSED
NOT STERILE
END TIME 3:18

Digital Signature #
B87A774B2A5ED7E0
