
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001636
16:45 29/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:11
PRESSURIZING 2:16
STERILIZING 3:14
136.5 C 224kPa 3:14
136.4 C 224kPa 3:44
136.4 C 222kPa 4:14
136.4 C 227kPa 4:44
136.4 C 225kPa 5:14
136.3 C 226kPa 5:44
136.2 C 223kPa 6:14
136.3 C 225kPa 6:44
136.3 C 225kPa 6:44
VENTING 6:44
AIR DRYING 7:38
CYCLE COMPLETE 7:44

Digital Signature #
489AE3DE46C1E93E
