
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 002901
13:13 29/12/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:36
STERILIZING 3:36
136.5 C 220kPa 3:36
136.4 C 224kPa 4:06
136.3 C 224kPa 4:36
136.4 C 225kPa 5:06
136.2 C 224kPa 5:36
136.3 C 222kPa 6:06
136.4 C 225kPa 6:36
135.9 C 217kPa 7:06
135.9 C 217kPa 7:06
VENTING 7:06
AIR DRYING 8:03
CYCLE COMPLETE 8:50

Digital Signature #
154E1B92C722AA33
