
Midmark M11 - v1.0.5
Total Cycles: 2183

Name: RITTER2
Sterilizer ID: V2313228
03/28/2023 07:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 80.3 F 0.0
2:00 81.2 F 0.0
4:00 87.9 F 0.0
6:00 114.1 F 0.0
8:00 148.7 F 3.0
10:00 175.9 F 6.6
12:00 199.2 F 10.6
14:00 213.1 F 6.5
16:00 223.1 F 8.3
18:00 234.5 F 12.6
20:00 233.9 F 9.2
22:00 239.8 F 10.4
24:00 244.7 F 13.1
26:00 256.3 F 19.2
28:00 266.7 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.4 F 29.5
2:00 271.2 F 29.1
3:00 272.3 F 29.5
4:00 271.1 F 29.0

Min 271.1 F 28.8
Max 272.5 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.6 F 0.0
5:00 203.5 F 0.0
10:00 201.1 F 0.0
15:00 204.6 F 0.0
20:00 207.0 F 0.0
25:00 217.5 F 0.0
30:00 211.4 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 29:33
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:07:34
CYCLE COMPLETE

03/28/2023 08:37 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____