

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1631

Name: RITTER 1  
Sterilizer ID: V2178198  
02/14/2022 03:30 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 117.2 F 0.0  
2:00 124.5 F 0.9  
4:00 157.1 F 3.5  
6:00 183.7 F 7.7  
8:00 202.3 F 12.7  
10:00 218.4 F 5.7  
12:00 230.5 F 10.6  
14:00 231.8 F 7.1  
16:00 243.9 F 13.1  
18:00 242.9 F 11.6  
20:00 255.9 F 18.3  
22:00 267.4 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.4 F 28.7  
2:00 271.7 F 28.5  
3:00 272.4 F 28.7  
4:00 271.5 F 28.4

Min 271.4 F 28.1  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.9 F 0.0  
5:00 230.5 F 0.0  
10:00 224.7 F 0.0  
15:00 228.0 F 0.0  
20:00 225.0 F 0.0  
25:00 226.1 F 0.0  
30:00 229.7 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 23:24  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:01:33  
CYCLE COMPLETE

02/14/2022 04:31 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_