
Midmark M11 - v1.0.0
Total Cycles: 5211

Name: RITTER 1
Sterilizer ID: V2178198
12/20/2024 12:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.6 F 0.0
2:00 167.3 F 0.0
4:00 174.7 F 0.1
6:00 199.6 F 0.9
8:00 221.0 F 4.8
10:00 224.4 F 8.7
12:00 234.2 F 12.9
14:00 235.8 F 9.2
16:00 241.4 F 10.6
18:00 245.7 F 12.8
20:00 243.5 F 11.7
22:00 255.7 F 17.9
24:00 266.1 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 27.9
1:00 271.4 F 27.9
2:00 271.9 F 28.3
3:00 272.5 F 28.7
4:00 272.7 F 28.7

Min 271.4 F 27.8
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.8 F 0.0
5:00 223.4 F 0.0
10:00 223.0 F 0.0
15:00 219.5 F 0.0
20:00 206.3 F 0.0
25:00 221.2 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 25:48
STERILIZING: 4:00
VENTING: 3:53
DRYING: 30:00
TOTAL CYCLE: 01:04:26
CYCLE COMPLETE

12/20/2024 01:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____