
Midmark M11 - v1.0.5
Total Cycles: 2796

Name: RITTER2
Sterilizer ID: V2313228
08/25/2023 09:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.4 F 0.0
2:00 172.6 F 0.0
4:00 179.7 F 2.0
6:00 199.5 F 6.1
8:00 215.9 F 10.4
10:00 228.6 F 7.0
12:00 234.3 F 11.1
14:00 234.7 F 8.8
16:00 240.2 F 10.5
18:00 247.1 F 14.1
20:00 258.6 F 20.4
22:00 268.9 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.9 F 29.7
2:00 271.8 F 29.0
3:00 271.5 F 29.3
4:00 272.6 F 29.7

Min 271.1 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.9 F 0.0
5:00 226.0 F 0.0
10:00 223.2 F 0.0
15:00 220.3 F 0.0
20:00 229.6 F 0.0
25:00 216.7 F 0.0
30:00 211.2 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 23:12
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:02:07
CYCLE COMPLETE

08/25/2023 11:00 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____