
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002152
14:56 16/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:53
PRESSURIZING 1:59
STERILIZING 3:04
136.5 C 225kPa 3:04
136.4 C 225kPa 3:34
136.0 C 223kPa 4:04
136.4 C 224kPa 4:34
136.4 C 225kPa 5:04
136.4 C 225kPa 5:34
136.4 C 225kPa 6:04
136.4 C 225kPa 6:34
136.4 C 225kPa 6:34
VENTING 6:34
AIR DRYING 7:25
CYCLE COMPLETE 37:25

Digital Signature #
98EE9046B64B7C44
