
Midmark M11 - v1.0.0
Total Cycles: 4282

Name: RITTER 1
Sterilizer ID: V2178198
02/07/2024 04:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.6 F 0.0
2:00 125.3 F 1.6
4:00 140.0 F 4.3
6:00 196.2 F 7.9
8:00 194.4 F 12.5
10:00 209.6 F 4.5
12:00 221.0 F 8.8
14:00 233.1 F 13.2
16:00 235.3 F 9.2
18:00 240.1 F 10.3
20:00 245.0 F 12.7
22:00 256.6 F 18.9
24:00 267.7 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.3 F 28.7
2:00 271.5 F 28.5
3:00 272.1 F 28.7
4:00 271.4 F 28.4

Min 271.2 F 28.2
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.3 F 0.0
5:00 231.2 F 0.0
10:00 229.9 F 0.0
15:00 230.7 F 0.0
20:00 228.1 F 0.0
25:00 229.0 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 25:18
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:03:56
CYCLE COMPLETE

02/07/2024 05:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____