

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4188

Name: RITTER 1  
Sterilizer ID: V2178198  
01/10/2024 12:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 159.6 F 0.0  
2:00 159.8 F 2.7  
4:00 180.1 F 6.0  
6:00 199.3 F 10.5  
8:00 216.3 F 5.5  
10:00 225.0 F 8.8  
12:00 235.0 F 12.8  
14:00 236.3 F 10.7  
16:00 235.5 F 8.8  
18:00 248.5 F 15.3  
20:00 259.9 F 21.6  
22:00 270.4 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.1  
1:00 272.5 F 29.9  
2:00 271.6 F 29.2  
3:00 271.4 F 29.6  
4:00 273.3 F 30.4

Min 271.0 F 29.0  
Max 273.3 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 216.3 F 0.0  
5:00 227.6 F 0.0  
10:00 226.5 F 0.0  
15:00 232.7 F 0.0  
20:00 231.2 F 0.0  
25:00 214.7 F 0.0  
30:00 213.5 F 0.0  
Drying Complete

FILLING: 1:20  
HEATING: 22:52  
STERILIZING: 4:00  
VENTING: 3:53  
DRYING: 30:00  
TOTAL CYCLE: 01:02:06  
CYCLE COMPLETE

01/10/2024 01:03 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_