

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001147  
10:45 29/06/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:12  
PRESSURIZING 2:24  
STERILIZING 3:47  
136.4 C 222kPa 3:47  
136.3 C 223kPa 4:17  
136.2 C 222kPa 4:47  
136.1 C 221kPa 5:17  
136.4 C 225kPa 5:47  
136.1 C 221kPa 6:17  
136.3 C 223kPa 6:47  
136.2 C 223kPa 7:17  
136.2 C 223kPa 7:17  
VENTING 7:17  
AIR DRYING 8:05  
CYCLE COMPLETE 23:05  
-----

Digital Signature #  
3FC1AA0CFA1E75C3  
-----