
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001294
12:14 13/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:49
PRESSURIZING 3:51
STERILIZING 5:04
136.0 C 221kPa 5:04
136.2 C 224kPa 5:34
136.4 C 225kPa 6:04
136.4 C 225kPa 6:34
136.4 C 224kPa 7:04
136.0 C 222kPa 7:34
136.4 C 225kPa 8:04
136.4 C 224kPa 8:34
136.4 C 224kPa 8:34
VENTING 8:34
AIR DRYING 9:29
CYCLE COMPLETE 13:09

Digital Signature #
F0EAA4B7A1DB7F77
