
Midmark M11 - v1.0.5
Total Cycles: 2414

Name: RITTER2
Sterilizer ID: V2313228
05/25/2023 10:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.4 F 0.0
2:00 161.2 F 3.1
4:00 176.0 F 7.2
6:00 197.5 F 12.2
8:00 214.9 F 6.7
10:00 230.0 F 12.2
12:00 234.1 F 9.5
14:00 236.7 F 9.1
16:00 248.2 F 14.9
18:00 260.3 F 21.7
20:00 271.1 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.5 F 29.6
2:00 271.9 F 29.8
3:00 272.2 F 29.4
4:00 272.0 F 29.9

Min 271.5 F 29.1
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.9 F 0.0
5:00 224.0 F 0.0
10:00 228.1 F 0.0
15:00 228.9 F 0.0
20:00 226.3 F 0.0
25:00 225.8 F 0.0
30:00 224.6 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 20:39
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:03
CYCLE COMPLETE

05/25/2023 11:06 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____