
Midmark M11 - v1.0.0
Total Cycles: 688

Name: MIDMARK
Sterilizer ID: V2178198
02/22/2021 03:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.3 F 0.0
2:00 150.2 F 2.5
4:00 172.9 F 6.5
6:00 204.4 F 12.4
8:00 223.0 F 7.5
10:00 235.0 F 13.1
12:00 240.7 F 11.1
14:00 238.0 F 9.5
16:00 253.3 F 16.9
18:00 265.9 F 24.4
20:00 272.3 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 271.8 F 28.7
2:00 272.2 F 29.1
3:00 272.3 F 29.2
4:00 271.0 F 28.6

Min 271.0 F 28.5
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.9 F 0.0
5:00 205.3 F 0.0
10:00 210.2 F 0.0
15:00 211.8 F 0.0
20:00 210.4 F 0.0
25:00 210.0 F 0.0
30:00 207.5 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 20:09
STERILIZING: 4:00
VENTING: 2:52
DRYING: 30:00
TOTAL CYCLE: 00:58:18
CYCLE COMPLETE

02/22/2021 04:29 PM
ACCEPTED: X
REJECTED:
Operator: U01