
Midmark M11 - v1.0.5
Total Cycles: 3635

Name: RITTER2
Sterilizer ID: V2313228
05/22/2024 12:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 91.2 F 0.0
2:00 91.7 F 0.1
4:00 98.6 F 1.1
6:00 125.7 F 3.5
8:00 161.9 F 6.4
10:00 181.2 F 9.5
12:00 197.5 F 13.0
14:00 209.0 F 5.3
16:00 218.6 F 8.8
18:00 229.4 F 12.7
20:00 231.8 F 9.1
22:00 238.7 F 11.6
24:00 242.1 F 12.3
26:00 253.0 F 17.9
28:00 263.0 F 23.8
30:00 272.2 F 30.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 30.0
1:00 272.1 F 29.9
2:00 272.6 F 30.2
3:00 272.1 F 30.5
4:00 271.7 F 29.9

Min 271.6 F 29.8
Max 272.6 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.1 F 0.0
5:00 217.5 F 0.0
10:00 209.0 F 0.0
15:00 194.5 F 0.0
20:00 192.2 F 0.0
25:00 194.8 F 0.0
30:00 197.0 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 31:15
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:09:50
CYCLE COMPLETE

05/22/2024 02:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____