

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4157

Name: RITTER 1  
Sterilizer ID: V2178198  
12/27/2023 10:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 187.0 F 0.0  
2:00 172.2 F 2.5  
4:00 173.9 F 6.1  
6:00 206.8 F 10.8  
8:00 218.9 F 4.6  
10:00 233.1 F 9.8  
12:00 233.4 F 8.6  
14:00 243.4 F 12.3  
16:00 241.3 F 10.8  
18:00 254.4 F 17.5  
20:00 266.0 F 24.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.1  
1:00 272.4 F 28.9  
2:00 271.6 F 28.1  
3:00 272.4 F 28.9  
4:00 271.9 F 28.3

Min 271.2 F 28.0  
Max 272.6 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.4 F 0.0  
5:00 243.5 F 0.0  
10:00 234.0 F 0.0  
15:00 228.7 F 0.0  
20:00 227.1 F 0.0  
25:00 230.3 F 0.0  
30:00 229.2 F 0.0  
Drying Complete

FILLING: 1:21  
HEATING: 21:28  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:00:11  
CYCLE COMPLETE

12/27/2023 11:44 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_