
Midmark M11 - v1.0.0
Total Cycles: 4127

Name: RITTER 1
Sterilizer ID: V2178198
12/15/2023 03:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.8 F 0.0
2:00 187.3 F 4.2
4:00 208.6 F 9.4
6:00 223.8 F 6.9
8:00 234.8 F 10.0
10:00 234.6 F 8.1
12:00 245.5 F 12.5
14:00 245.7 F 12.8
16:00 258.9 F 19.9
18:00 270.5 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.4 F 28.4
2:00 272.1 F 28.5
3:00 272.3 F 28.2
4:00 272.1 F 28.5

Min 271.5 F 27.8
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.2 F 0.0
5:00 229.5 F 0.0
10:00 200.5 F 0.0
15:00 193.9 F 0.0
20:00 197.0 F 0.0
25:00 186.5 F 0.0
30:00 184.4 F 0.0
Drying Complete

FILLING: 2:39
HEATING: 18:46
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 00:58:46
CYCLE COMPLETE

12/15/2023 04:51 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____