
Midmark M11 - v1.0.0
Total Cycles: 2700

Name: RITTER 1
Sterilizer ID: V2178198
12/20/2022 02:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.8 F 0.0
2:00 198.4 F 4.2
4:00 203.9 F 9.6
6:00 221.0 F 6.2
8:00 235.6 F 11.2
10:00 237.2 F 9.1
12:00 239.2 F 9.1
14:00 250.4 F 15.2
16:00 262.8 F 22.6
18:00 271.8 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.3 F 28.8
2:00 272.0 F 28.8
3:00 272.1 F 28.7
4:00 271.9 F 28.8

Min 271.5 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.0 F 0.0
5:00 222.7 F 0.0
10:00 232.3 F 0.0
15:00 236.1 F 0.0
Dry Cycle Aborted

FILLING: 1:49
HEATING: 18:05
STERILIZING: 4:00
VENTING: 3:12
DRYING: 17:47
TOTAL CYCLE: 00:44:52

12/20/2022 02:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____