
Midmark M11 - v1.0.0
Total Cycles: 5079

Name: RITTER 1
Sterilizer ID: V2178198
11/12/2024 09:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.5 F 0.0
2:00 71.4 F 0.0
4:00 82.1 F 0.0
6:00 112.9 F 0.1
8:00 145.9 F 0.9
10:00 173.1 F 0.9
12:00 195.5 F 3.2
14:00 210.2 F 6.2
16:00 215.7 F 9.9
18:00 227.6 F 11.8
20:00 231.6 F 8.2
22:00 241.4 F 12.9
24:00 239.5 F 10.2
26:00 236.5 F 8.2
28:00 249.4 F 14.6
30:00 260.7 F 20.9
32:00 271.0 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 27.9
1:00 272.4 F 28.4
2:00 272.0 F 28.4
3:00 272.0 F 28.1
4:00 272.5 F 28.7

Min 271.5 F 27.8
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.1 F 0.0
5:00 194.4 F 0.0
10:00 185.5 F 0.0
15:00 185.8 F 0.0
20:00 215.8 F 0.0
25:00 229.1 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 0:43
HEATING: 32:42
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:10:55
CYCLE COMPLETE

11/12/2024 10:38 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____