
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002666
8:41 11/11/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:01
PRESSURIZING 3:08
STERILIZING 4:16
136.4 C 220kPa 4:16
136.1 C 221kPa 4:46
136.4 C 221kPa 5:16
136.3 C 221kPa 5:46
136.2 C 222kPa 6:16
136.4 C 222kPa 6:46
136.2 C 223kPa 7:16
136.4 C 223kPa 7:46
136.4 C 223kPa 7:46
VENTING 7:46
AIR DRYING 8:36
CYCLE COMPLETE 8:40

Digital Signature #
846F4C5DF88689C5
