
Midmark M11 - v1.0.0
Total Cycles: 2148

Name: RITTER 1
Sterilizer ID: V2178198
07/19/2022 08:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.9 F 0.0
2:00 129.4 F 0.0
4:00 157.9 F 0.0
6:00 177.9 F 3.4
8:00 188.9 F 8.1
10:00 204.7 F 12.4
12:00 217.0 F 5.8
14:00 228.5 F 10.0
16:00 236.9 F 8.9
18:00 240.2 F 11.9
20:00 237.8 F 9.8
22:00 250.8 F 16.2
24:00 262.0 F 22.6
26:00 272.5 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.6 F 29.4
2:00 271.4 F 28.7
3:00 272.6 F 29.6
4:00 271.9 F 28.9

Min 271.4 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.4 F 0.0
5:00 201.1 F 0.0
10:00 214.3 F 0.0
15:00 211.5 F 0.0
20:00 213.7 F 0.0
25:00 220.3 F 0.0
30:00 223.7 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 26:27
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:04:31
CYCLE COMPLETE

07/19/2022 09:45 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____