

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4339

Name: RITTER 1  
Sterilizer ID: V2178198  
02/27/2024 04:44 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 104.4 F 0.0  
2:00 110.6 F 0.9  
4:00 118.0 F 2.9  
6:00 145.5 F 5.9  
8:00 171.8 F 9.5  
10:00 201.7 F 10.2  
12:00 213.2 F 5.6  
14:00 223.9 F 9.5  
16:00 235.1 F 11.0  
18:00 235.8 F 9.9  
20:00 238.4 F 9.5  
22:00 245.0 F 13.0  
24:00 256.1 F 18.8  
26:00 266.5 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.5  
1:00 272.6 F 29.2  
2:00 271.4 F 28.4  
3:00 272.3 F 29.3  
4:00 272.2 F 28.9

Min 271.1 F 28.4  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.6 F 0.0  
5:00 210.2 F 0.0  
10:00 215.6 F 0.0  
15:00 223.8 F 0.0  
20:00 235.1 F 0.0  
25:00 229.5 F 0.0  
30:00 229.6 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 27:39  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 01:06:13  
CYCLE COMPLETE

02/27/2024 05:50 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_