

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3734

Name: RITTER2  
Sterilizer ID: V2313228  
06/26/2024 08:22 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.1 F 0.0  
2:00 80.8 F 0.0  
4:00 76.2 F 0.1  
6:00 119.2 F 1.2  
8:00 146.5 F 4.2  
10:00 176.6 F 7.3  
12:00 198.4 F 11.1  
14:00 215.0 F 4.3  
16:00 224.9 F 7.1  
18:00 234.8 F 10.8  
20:00 229.7 F 7.5  
22:00 240.5 F 11.7  
24:00 237.6 F 9.7  
26:00 250.4 F 15.9  
28:00 261.1 F 22.2  
30:00 271.5 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.9  
1:00 272.8 F 29.7  
2:00 271.6 F 28.8  
3:00 272.6 F 29.9  
4:00 272.3 F 29.1  
  
Min 271.3 F 28.8  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.5 F 0.0  
5:00 221.5 F 0.0  
10:00 193.0 F 0.0  
15:00 212.6 F 0.0  
20:00 208.7 F 0.0  
25:00 204.0 F 0.0  
30:00 199.0 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 30:40  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 01:08:47  
CYCLE COMPLETE

06/26/2024 09:30 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_