
Midmark M11 - v1.0.0
Total Cycles: 5498

Name:
Sterilizer ID: V2178198
04/11/2025 10:42 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 109.6 F 0.0
2:00 117.4 F 0.8
4:00 122.9 F 3.0
6:00 145.4 F 5.6
8:00 169.0 F 8.6
10:00 190.0 F 12.0
12:00 207.8 F 2.7
14:00 212.7 F 6.4
16:00 223.6 F 10.1
18:00 236.5 F 9.2
20:00 238.7 F 10.2
22:00 238.2 F 9.1
24:00 246.2 F 13.1
26:00 256.6 F 18.7
28:00 265.9 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.2
1:00 271.8 F 29.0
2:00 272.4 F 29.1
3:00 271.4 F 28.7

Min 270.9 F 28.1
Max 273.1 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.2 F 0.0
5:00 189.2 F 0.0
10:00 173.8 F 0.0
15:00 172.1 F 0.0
20:00 173.6 F 0.0
25:00 181.8 F 0.0
30:00 196.4 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 29:54
STERILIZING: 3:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:07:34
CYCLE COMPLETE

04/11/2025 11:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____